

## **A quasi-monochromatic electron beam of "PAHRA" accelerator for calibration of detectors**

*Tuesday, 23 October 2018 10:00 (15)*

The characteristics of the calibration quasi-monochromatic beam of secondary electrons of the accelerator S-25R "PAHRA" of the P. N. Lebedev Physics Institute of Russian Academy of Sciences based on the magnet SP-57 are presented. The energy resolution of the beam from a copper converter in the thickness range of 0.1 - 5 mm and the interpolar gap of the magnet 6 cm in the energy range of the electron beam  $E = 98 - 294$  MeV amounted to  $\delta = 10 - 4.5$  %, respectively.

**Primary author(s)** : Dr. BASKOV, Vladimir (Lebedev Physical Institute RAS)

**Presenter(s)** : Dr. BASKOV, Vladimir (Lebedev Physical Institute RAS)

**Session Classification** : Facilities and Advanced Detector Technologies

**Track Classification** : Facilities and advanced detector technologies