

Xenon scintillation response in two-phase emission detector

Monday, 10 October 2016 16:39 (13)

The analysis of experimental data of the two-phase xenon emission detector has been conducted. There were determined ranges of parameters for neutrons and gamma events, using a two-dimensional diagram in isolines. The original method for estimating the deexcitation time of liquid xenon as a working environment of the detector was applied.

Primary author(s) : Mr. BURENKOV, Alexander (ITEP, MEPhI); Mr. KONOVALOV, Alexey (ITEP, MEPhI, MIPT); Mr. LUKYASHIN, Anton (ITEP, MEPhI); Dr. AKIMOV, Dmitry (ITEP, MEPhI); Mr. BELOV, Vladimir (ITEP, MEPhI)

Presenter(s) : Mr. LUKYASHIN, Anton (ITEP, MEPhI)

Session Classification : Methods of experimental physics - parallel I

Track Classification : Methods of experimental physics